

rotronic HYGROGEN

HygroClip IC 60862521 AutoCal Certificate

Date (YYYY-MM-DD): 2013-07-18
 Time (HH:MM:SS): 23:21:30

Probe Under Test: HygroClip IC
 Serial No.: 60862521

Reference: HygroClip2
 Serial No.: 60501780
 Reference Calibration Standard:
 HC2-S
 Serial No.: 60501780
 Reference Last Calibration Date: 2013-02-25
 Reference Next Calibration Date: 2014-02-25
 Comment on Reference:
 Certificate no. of reference probe
 SZ20131227
 Swiss Calibration Service
 3T-4H
 HygroGen2 ID: 912772.79

AutoCal Program Number: 4
 Descriptor: ACal Profile 4
 Autocal Program Structure:

Step:	Temperature SP (°C):	Humidity SP (%):	Step Duration (HH:MM):	Adjust Temperature:	Adjust Humidity:
1	9.0	50.0	03:00	No	No
2	19.0	50.0	03:00	No	No
3	29.0	35.0	03:00	No	No

Tolerance to Set Point at which step duration timing will occur (zero = no tolerance applied).
 SP Temperature Tolerance (°C): 0.5
 SP Relative Humidity Tolerance (%): 1.0

Probe Actions Recorded:

Step:	Action:	Serial No.:	Date:	Time:	Probe Temp. (°C):	Probe RH (%):	Ref. Temp. (°C):	Ref. RH (%):	Temp. Correction:	RH Correction:
1	Read	60862521	2013-07-18	17:06:55	8.94	49.28	9.00	49.98	0.06	0.70
1	Read	60862521	2013-07-18	17:08:55	8.93	49.33	9.00	49.99	0.07	0.66
1	Read	60862521	2013-07-18	17:10:55	8.94	49.33	9.00	50.03	0.06	0.70
1	Read	60862521	2013-07-18	17:12:54	8.91	49.33	9.00	50.03	0.09	0.70

Step:	Action:	Serial No.:	Date:	Time:	Probe Temp. (°C):	Probe RH (%):	Ref. Temp. (°C):	Ref. RH (%):	Temp. Correction:	RH Correction:
1	Read	60862521	2013-07-18	17:14:24	8.93	49.33	9.00	50.01	0.07	0.68
2	Read	60862521	2013-07-18	20:10:52	18.89	49.32	19.00	49.99	0.11	0.67
2	Read	60862521	2013-07-18	20:12:52	18.89	49.32	19.00	49.99	0.11	0.67
2	Read	60862521	2013-07-18	20:14:51	18.89	49.34	19.00	50.01	0.11	0.67
2	Read	60862521	2013-07-18	20:16:51	18.88	49.32	19.00	49.99	0.12	0.67
2	Read	60862521	2013-07-18	20:18:21	18.90	49.37	19.00	50.04	0.10	0.67
3	Read	60862521	2013-07-18	23:13:25	28.85	34.15	29.00	35.01	0.15	0.86
3	Read	60862521	2013-07-18	23:15:25	28.84	34.15	29.00	34.99	0.16	0.84
3	Read	60862521	2013-07-18	23:17:25	28.84	34.15	29.00	35.01	0.16	0.86
3	Read	60862521	2013-07-18	23:19:24	28.84	34.15	29.00	34.99	0.16	0.84
3	Read	60862521	2013-07-18	23:20:54	28.85	34.15	29.00	34.99	0.15	0.84

Operator: nbp

Operator Signature: _____ Date: _____

Operator Comment:

Periode afsluttet

Contact Details:

C.K. Environment A/S

Walgerholm 3

3500 Vaerlose

Telephone: 45 44989906

Email: cke@cke.dk